Name of Document	M/S Bio Madics	M/S Gulf Marketing	M/S Martin Dow Pvt. Ltd	M/S Popular International	M/S A.H Traders	M/S A.M Healthcare	M/S Diagnostic Medical Associate	M/S Star Traders
Tender Purchase Receipt Attached	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
NIC attested copy attached	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
NTN attested copy attached	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
STN attested copy attached	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
One Year Business History	Not Relevant	NO.	Yes	NO.	NO.	Yes	Yes	NO.
Authority Letter / distribution certificate of manufacture / Importer	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Establishment License/ License to import	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Name of Manufacturer of instrument	Austria	USA	Sweden	USA	Turkey	Japan	Hungry	Germany
Model Name and Model No.	Cellagon-3	Mythic TM 18	Swelab Alfa Plus	Mission HA 360	BT Pro-021	Sysmax XP-100	Abacus 380	Erba H-360
Country of origin	Hungry	USA	Sweden	USA	Turkey	Japan	Hungry	Germany
Country of manufacturing of	Hungry	USA	Sweden	USA	Turkey	Japan	Hungry	Germany

Name of Document	M/S Bio Madics	M/S Gulf Marketing	M/S Martin Dow Pvt. Ltd	M/S Popular International	M/S A.H Traders	M/S A.M Healthcare	M/S Diagnostic Medical Associate	M/S Star Traders
instrument								
Quality Certificate	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Affidavit regarding non black listing of the firm	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Affidavit regarding Price Reasonability Certificate of the firm	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Technically Eligibility of the firm	Not Eligible	Not Eligible	Eligible	Not Eligible	Not Eligible	Eligible	Eligible	Not Eligible

	M/S Bio Madics	M/S Gulf Marketing	M/S Martin Dow Pvt. Ltd	M/S Popular International	M/S A.H Traders	M/S A.M Healthcare	M/S Diagnostic Medical Associate	M/S Star Traders
Detailed Requirements: Automated hematology counter Auto probe cleaning/wiping RS232/USB interface 220V, 50 Hz operated	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
User Adjustable Settings: Through put 60 samples/hour or more 3- Parts differential with minimum 20 parameters Histograms for WBC RBC and platelet Sample volume Max 120ul of whole blood		Yes	Yes	Yes	Yes	Yes	Yes	Yes

			INICAL EVALUATIO	J. CO	(E) OIII			
or less								
Auto sampling with								
Loading Capacity of								
30 to 50 Samples								
Displayed	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Parameters:								
LED displayed with								
touch screen								
Auto calibration								
programme								
Patient data entry and								
specimen recognition								
by keyboard/LED								
touch screen/Bar Code								
Code								
Patient data and								
results storage of 5,000 results or more								
Accessories:	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
External laser printer								
Consumables,								
reagents, calibrations,								
controls and start up								

			INICAL EVALUATION			1		
(free of cost)								
Compatible imported								
online sine wave UPS								
with battery backup								
for 30 Minutes								
(Emerson, Liebert,								
Chloride, MGE, APC								
or equivalent)								
Operation Manual								
with a soft copy								
Service manual with a								
soft copy								
One Backup model								
with same								
specification								
AC for temperature								
maintenance								
Free of cost parts								
replacement, services								
and maintenance &								
visits of engineer	35/074 35 34	7.5/0.00 7.0	35/035 11 3	15/07	7.5/0 / 77	7.5/2 / 7.5	7.570	3.510.0
	M/S Bio Madics	M/S Gulf	M/S Martin Dow Pvt. Ltd	M/S Popular International	M/S A.H Traders	M/S A.M Healthcare	M/S Diagnostic	M/S Star Traders
		Marketing	rvi. Liu	mernauonai	Trauers	пеанисаге	Medical	Trauers
							Associate	
Complete Blood	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Count (CBC)								

DECISION	OF	Not Eligible	Not Eligible	Eligible	Not Eligible	Not Eligible	Eligible	Eligible	Not Eligible
TECHNICAL									
ADVISORY									
COMMITTEE									