

TENDER FOR THE PLACEMENT OF ELISA ANALYZER ON REAGENT RENTAL BASIS FOR THE YEAR 2022-23

TECHNICAL EVALUATION COMMITTEE REPORT

| Name of Document | M/S Diagnostic Medical Associate Rawalpindi |
|---|--|
| Tender Purchase Receipt Attached | Yes |
| NIC attested copy attached | Yes |
| NTN attested copy attached | Yes |
| STN attested copy attached | Yes |
| Country of manufacture | Yes |
| Country of origin of equipment, Brand Name, Model No. | Yes |
| Country of Origin of Kit/ Packs, Catalogue No. Brand Name | Yes |
| Availability of model on official website | Yes |
| Valid Quality Certificate (ISO 13485:2003 / CE Marked / FDA/ MHLW/ IVD) | Yes |
| Valid Establishment License/ License to import | Yes |
| Valid Authorization Certificate from Manufacturer / foreigner principal/ Sole Agent | Yes |
| Registration certificate with DRAP (if applicable) | Yes |
| Affidavit that the firm was / is not black listed from any Government institution | Yes |
| Affidavit regarding price reasonability certificate | Yes |
| One year relevant business history | Yes |
| Technically Eligibility of the firm | Eligible |

TENDER FOR THE PLACEMENT OF ELISA ANALYZER ON REAGENT RENTAL BASIS FOR THE YEAR 2022-23

TECHNICAL EVALUATION COMMITTEE REPORT

| TECHNICAL SPECIFICATIONS:- | M/S Diagnostic Medical Associate Rawalpindi |
|--|--|
| <p>Detailed Requirements Open system fully automated microprocessor controlled Microplate/ Batch ELISA Analyzer with LCD monitor and data monitoring system. Automatic sample identification dilution, incubation, conjugation, washing and reading facility. Programmable for 04 plates of 96 wells with filters of 405-750 nm with at least 06 filters. Throughput 4x96 samples test. Auto clot detection. 8 or more suction channels and at least 8 dispensing channels. System should have possibility to run 25 or more controls. Digital display for results. External laser printer. Imported, compatible since UPS for backup of 60 minutes.</p> | <p>OK</p> |
| <p>Final Bidder of the tender will provide the distillation plant for ELISA Section. Final Bidder of the tender will provide all consumables and accessories.</p> <p>All the participants of tender will place the instrument at the time of kits evaluation at Pathology Department QAMC. Final Bidder of the tender will provide an instrument for backup</p> | <p>OK</p> |

TENDER FOR THE PLACEMENT OF ELISA ANALYZER ON REAGENT RENTAL BASIS FOR THE YEAR 2022-23

TECHNICAL EVALUATION COMMITTEE REPORT

| TECHNICAL SPECIFICATIONS:- | M/S Diagnostic Medical Associate Rawalpindi |
|-----------------------------------|--|
| Name of Test | |
| Anti HCV ELISA Kits | OK |
| HBsAg ELISA Kits | OK |
| HEV Ag ELISA | OK |
| Anti HIV ELISA | OK |
| Decision for TAC | Eligible |