

BRAND NEW BLOOD CULTURE SYSTEM ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24

TECHNICAL EVALUATION COMMITTEE REPORT

| | |
|--|--|
| Name of Document | M/S Global Markeeting service Rawalpindi |
| Tender Purchase Receipt Attached | Yes |
| NIC attested copy attached | Yes |
| NTN attested copy attached | Yes |
| STN attested copy attached | Yes |
| One Year Business History | Yes |
| Authority Letter / distribution certificate of manufacture / Importer | Yes |
| Establishment License/ License to import | Yes |
| Name of Manufacturer of instrument | BIOMEREUX |
| Model Name and Model No. | Blood Culture System Model:- BacT Alert 3D120 |
| Country of origin | USA |
| Country of manufacturing of instrument | France |
| Quality Certificate | Yes |
| Affidavit regarding non black listing of the firm | Yes |
| Affidavit regarding Price Reasonability Certificate of the firm | Yes |
| Technically Eligibility of the firm | Eligible |

BRAND NEW BLOOD CULTURE SYSTEM ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24
TECHNICAL EVALUATION COMMITTEE REPORT

| | |
|--|--|
| | M/S Global Marketing service Rawalpindi |
| <p><u>Detailed Requirement:</u></p> <p>Fully automated testing</p> <p>System will be non-radiometric</p> <p>Continuous- monitoring instruments</p> <p>Automated quality control to ensure precise and reliable operation</p> <p>Advanced algorithms for rapid detection of pathogens in blood culture like individual bottle types, like low blood volume, pediatric specimens, or to detect slow growing organisms</p> <p>RS232 or USB interface</p> <p>220V 50 Hz, AC</p> | OK |
| <p><u>User Adjustable Settings:</u></p> <p>Reagent bottles will be tested every 10 minutes</p> <p>Internal temperature on 70°F to 85°F</p> <p>Average capability of approximately 20-25 specimens per day</p> <p>Total capacity at least 100</p> <p>Data management systems with bar-code scanning</p> <p>Long term data storage and historical trending with bi- directional interface to an LIS</p> | OK |

BRAND NEW BLOOD CULTURE SYSTEM ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24

TECHNICAL EVALUATION COMMITTEE REPORT

| | |
|---|-----------|
| <p><u>Displayed Parameters:</u></p> <p>Digital display for results</p> <p>Results will be displayed as positive/negative and numerical growth units</p> | <p>OK</p> |
| <p><u>Accessories:</u></p> <p>Complete with standard and operation accessories</p> <ul style="list-style-type: none">• Built-in or External laser Printer• Imported compatible sine wave UPS for back up of up to 30 minutes (Emerson, liebert, Chloride, MGE, APC or Equivalent)• Opening Manual with a soft copy• Service manual with a soft copy | <p>OK</p> |

BRAND NEW BLOOD CULTURE SYSTEM ON REAGENT RENTAL BASIS FOR THE YEAR 2023-24

TECHNICAL EVALUATION COMMITTEE REPORT

| | |
|---|---|
| Name of Tests | M/S Global Markeeting service Rawalpindi |
| Peads Blood Culture Bottles (With automated detection systems) | OK |
| Adults Blood Culture Bottles (With automated detection systems) | OK |
| Decision of Technical Advisory committee | Eligible |